



# Certificate of Traceability

## Manufacturer Certification

This certifies that **Boditech Med Inc.**, using **Immunoassay** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: <b>i-Chroma Reader</b>	Calibrator Lot: <b>130013A, 120688A</b>
Reagent Lot: <b>AA YXA39, AA YXA40, AA YXA41</b>	Calibrator Assigned Values: <b>5.8%, 9.2%</b>

Date of Certification: January 1, 2014

Certification Expires: January 1, 2015

*Daniel South*

NGSP Steering Committee Chair

*Randa R Little PhD.*

NGSP Network Coordinator

*Massa Umehara, Ph.D.*

SRL director/ supervisor



# Scatter Plot

Boditech Med Inc. Immunoassay on i-Chroma Reader

