## Monitor Veterinario 1.0 X12 VET







X12 VET Specification	1				
Physical Specifications	i				
Dimension	306±2 mm (W) × 309±2 mm (H) × 151±2 mm (D)				
Mary Mainlet	< 3.5 kg				
Max Weight	Standard confi	Standard configurations, no battery or accessories			
Power Supply					
Line Voltage	100 V to 240 V	<b>'~</b>			
Current	1.4 A to 0.7 A	77			
Frequency	50 Hz/60 Hz				
Battery					
Capacity	2550 mAh , 51	00 mAh			
On and the a Time	2550 mAh	≥ 4 h			
Operating Time	5100 mAh	≥ 8 h			
Observe Time	2550 mAh	≤ 3.5 h, 90% charge			
Charge Time	5100 mAh	≤ 6.5 h, 90% charge			
Display					
Display screen	12.1-inch color	TFT screen, touch scr	een available		
Resolution	800 × 600				
Waves	A maximum of 13 waveforms can be displayed on the same screen				
Recorder					
Record Width	48 mm				
Record Paper Width	50 mm				
Paper Speed	12.5 mm/s, 25 mm/s, 50 mm/s				
Channels	3				
	Continuous rea	al-time recording			
	8-second real-	time recording			
	20-second real	-time recording			
	Time recording  Alarm recording  Trend graph recording				
Recording Types	Trend table recording				
	NIBP review recording				
	Arrhythmia review recording				
	Alarm review recording				
	12-lead analysis recording				
	ST view recording				
	QT view recording				
Data Storage					
Internal Temporary			3 hrs, at 1 s resolution		
Memory	Trend graph/tre	end table review	120 hrs, at 1 min resolution		



	Alarm/Monitoring Event data	Up to 200 sets			
	NIBP Measurement Review	1200 sets			
Arrhythmia events		Up to 200 sets			
	12-lead Diagnosis Review	Up to 50 sets			
	single piece of patient data maximally contains the following information:				
	Trend graph and trend table	240 hours, at 1 min resolution			
Non-volatile Memory	NIBP measurement review	1200 sets			
(internal or external	Alarm review	200 sets			
storage device)	Arrhythmia event	200 sets			
	12-lead diagnosis review	50 sets			
	Waveforms	48 hours			
Wi-Fi					
IEEE	802.11a/b/g/n				
Frequency Band	2.4 GHz ISM band & 5 GHz ISM band				
Interfaces and others					
VGA output (optional)	///////////////////////////////////////	1			
USB interface		2			
Nurse call / analog output/ defibrillator synchronization (optional) 1					
Network Interface	Network Interface 1				
Data Transmission					
Data Export	Ethernet / USB				
ECG					
	3-Electrode: I, II, III				
Lead Mode	5-Electrode: I, II, III, aVR, aVL, aVF, V				
Edd Wode	6-Electrode: I, II, III, aVR, aVL, aVF, and leads responding to Va, Vb				
//////////	10-Elctrode: I, II, III, aVR, aVL, aVF	F, V1, V2, V3, V4, V5, V6			
Lead naming style	AHA, IEC				
Display Sensitivity	1.25 mm/mV (×0.125), 2.5 mm/mV (×0.25), 5 mm/mV (×0.5),10 mm/mV (×1),				
(Gain Selection)	20 mm/mV (×2), 40 mm/mV (×4), AUTO gain				
Sweep speed	6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s				
	Diagnosis: 0.05 Hz to 150 Hz				
	Diagnosis1: 0.05 Hz to 40 Hz				
Randwidth ( 2 dP)	Monitor: 0.5 Hz to 40 Hz				
Bandwidth (-3 dB)	Surgery: 1 Hz to 20 Hz				
	Enhanced: 2 Hz ~18 Hz				
	Customized: High-pass Filter and Low-pass Filter				



CMRR	Diagnosis: > 95 dB  Monitor: > 105 dB  Surgery: > 105 dB  Enhanced: > 105 dB  Diagnosis 1: > 105 dB (when Notch is turned on)  Customized: > 105 dB (Low-pass Filter < 40 Hz)  > 95 dB (Low-pass Filter > 40 Hz)				
Hum Filter	In diagnosis, diagnosis 1, monitor, surgery, enhanced modes: 50Hz/60 Hz (Hum filter can be turned on or off manually)				
Recovery time after defibrilla	tion <5 s				
ESU Protection	Cut mode: 300 W Coagulation mode: 100 W Restore time: ≤10 s				
Pace pulse detecting lead	one among I, II, III, AVR, A	VL, AVF, V1, V2, V3,V4, V5	i, V6		
Heart Rate					
Range	15 bpm to 300 bpm				
Accuracy	±1% or ±1 bpm, whicheve	r is greater			
Resolution	1 bpm				
PVC					
Range	0 to 350 PVCs/ min				
Resolution	1 PVCs/min				
ST value					
Range	-2.0 mV to +2.0 mV				
Accuracy	±0.02 mV or 10% (-0.8 mV to +0.8 mV), whichever is greater.  Beyond this range: not specified.				
Resolution	0.01 mV				
QT measurement					
Range	200 ms to 800 ms				
Resolution	4 ms				
Accuracy	±30 ms				
QTc measurement					
Range	200 ms to 800 ms				
Resolution	1 ms				
ΔQTc measurement					
Range	-600 ms to 600 ms				
Resolution	1 ms				
Arrhythmia analyses					
Asystole	Sustain VT	V-Fib/V-Tach	ExtremeTachy		



ExtremeBrady	V-Tach	Vent Brady	Tachy	
Brady	Wide QRS Tachy	Non-Sustain VT	Afib	
Vent Rhythm	Acc. Vent Rhythm	Pause	Pauses/min High	
PVCs High	R on T	PVC Bigeminy	PVC Trigeminy	
Pacer not Pacing	Pacer not Capture	Missed Beat	VEB	
PVC	Couplet	Run PVCs	Multiform PVCs	
IPVC	Irr Rhythm	PAC Bigeminy	PAC Trigeminy	
Low Voltage(Limb)				
12-lead ECG synchroni	zation analysis	·		
Average parameters of he	art beat	17	-313144/1/	
Heart rate (bpm)		1//1		
Time limit of P wave (ms)		3/////		
PR interval (ms)		2/2/2/2		
QRS interval (ms)		2////////		
QT/QTC (ms)				
P-QRS-T AXIS				
RESP				
Method	Trans-thoracic imped	lance: R-F(RA-LL), R-L (RA-	LA)	
Measurement lead	Options are lead I an	Options are lead I and II		
Measuring Range	0 rpm to 150 rpm			
Resolution	1 rpm			
Aggurgay	6 rpm to 150 rpm: ±2 rpm			
Accuracy	0 rpm to 5 rpm: not specified			
Gain Selection	×0.25, ×0.5, ×1, ×2,	×0.25, ×0.5, ×1, ×2, ×3, ×4, ×5		
Sweep	6.25 mm/s, 12.5 mm/s	6.25 mm/s, 12.5 mm/s, 25.0 mm/s, 50.0 mm/s		
Apnea Alarm Time	10 s, 15 s, 20 s, 25 s	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s		
NIBP	·			
Method	Oscillometry	Oscillometry		
Mode	Manual, Auto, Contin	Manual, Auto, Continuous		
Measuring Interval in Auto	1/2/3/4/5/10/15/30/60	111111111111111111111111111111111111111		
Mode	1/2/3/4/5/10/15/30/60/90/120/180/240/360/480 min			
Continuous	5 min, interval is 5 s	5 min, interval is 5 s		
Measuring Type	SYS, DIA, MAP, PR	SYS, DIA, MAP, PR		
		SYS: 25 mmHg to 290 mi	SYS: 25 mmHg to 290 mmHg	
	Big Cuff Mode	DIA: 10 mmHg to 250 mmHg		
		MAP: 15 mmHg to 260 mmHg		
Measuring Range		SYS: 25 mmHg to 240 mi	mHa.	
	The second secon	LOTO, ZO IIIIIII II IU Z4U III		
	Middle Cuff Mode	DIA: 10 mmHg to 200 mm		



			SYS: 25 mmHg to 140 mmHg
	Small Cuff Mode		DIA: 10 mmHg to 115 mmHg
			MAP: 15 mmHg to 125 mmHg
Cuff Pressure Measuring Range	0 mmHg to 300 mmHg		
Pressure Resolution	1 mmHg		
Maximum Mean Error	±5 mmHg		
Maximum Standard Deviation	8 mmHg		
Maximum Measuring	Big/Middle cuff	120	s
Period	Small cuff	90 s	
Typical Measuring Period	20 s to 35 s (dep	end o	n HR/motion disturbance)
	Big cuff	297	mmHg±3 mmHg
Overpressure Protection	Middle cuff	245	mmHg±3 mmHg
	Small cuff	147	mmHg±3 mmHg
SpO <sub>2</sub>			
Measuring Range	0% to 100%		
Resolution	1%		
Data update period	1 s		
Accuracy	±3% (70% to 100% SpO <sub>2</sub> )		
DI (Dorfusion Indox)	Undefined (0% to 69% SpO <sub>2</sub> )		
PI (Perfusion Index)	0.40		
Measuring Range Resolution	0-10		
TEMP	1		
11.4-14.11	2		
Channel	2		
Position	Skin, Oral cavity, Rectum		
Sensor type	YSI-10K and YSI-2.252K		
Technique	Thermal resistance		
Measure Parameter	T1, T2, TD		
Position	Skin, Oral, Rectum		
Unit	°C, °F		
Measuring Range	0°C to 50°C (32 °F to 122 °F)		
Resolution	0.1°C (0.1°F)		
Accuracy	Accuracy (not including sensor): ±0.1°C  Sensor accuracy: ≤ ±0.2°C		
Transient Response Time	≤ 30 s		
PR			
PR (SpO <sub>2</sub> )			4-1-11-1-11-1-1-1-1-1-1-1-1-1-1-1-1-1-1
	Measuring range		25 bpm to 300 bpm



	Measuring range	40 bpm to 240 bpm		
PR (NIBP)	Accuracy		reater	
	<u> </u>	±3 bpm or 3.5%, whichever is greater		
DD (IDD)	Measuring range	20 bpm to 300 bpm		
PR (IBP)	Accuracy	30 bpm to 300 bpm: ±2 bpm or 2	2%, whichever is greater	
IDD		20 bpm to 29 bpm: undefined		
IBP Change	1/0			
Channel	1/2			
Technique	Direct invasive measu	T.		
	Art	0 mmHg to +300 mmHg		
Measuring range	PA	-6 mmHg to +120mmHg		
	CVP/RAP/LAP/ICP	-10 mmHg to +40 mmHg		
56	P1/P2	-50 mmHg to +300 mmHg		
Resolution	1 mmHg			
±2% or ±1 mmHg, whichever is greater				
Accuracy	(not including sensor)			
Unit	kPa, mmHg, cmH2O	kPa, mmHg, cmH2O		
CO <sub>2</sub>	A))			
Measure Parameters	EtCO <sub>2</sub> , FiCO <sub>2</sub> , AwRI	EtCO <sub>2</sub> , FiCO <sub>2</sub> , AwRR		
Unit	mmHg, %, kPa		3/3/4///	
	EtCO <sub>2</sub>	0 mmHg to 150 mmHg (0% to 20%)		
Measuring Range	FiCO <sub>2</sub>	0 mmHg to 50 mmHg		
	AwRR	2 rpm to 150 rpm		
11/////	EtCO <sub>2</sub>	1 mmHg		
Resolution	FiCO <sub>2</sub>	1 mmHg		
	AwRR	1 rpm		
MARKER	111111111	±2 mmHg, 0 mmHg to 40	Typical conditions:	
		mmHg	Ambient temperature:	
		±5% of reading, 41 mmHg to	(25 ± 3) ° C	
		70 mmHg	Barometric pressure:	
	EtCO <sub>2</sub>	±8% of reading, 71 mmHg to	(760 ± 10) mmHg	
Accuracy		100 mmHg	Balance gas: N <sub>2</sub>	
Accuracy		±10% of reading, 101 mmHg	Sample gas flowrate:	
		to 150 mmHg	100 ml/min	
		±12% of reading or ±4 mmHg,		
		whichever is greater	All conditions	
	A 55			
	AwRR	±1 rpm		
Sample Gas Flowrate		70 ml/min or 100 ml/min(default), accuracy: ±15 ml/min		
Warm-up time	Display waveform within 20 s, Reach the design accuracy within 2 minutes.			
Response time	<4 s			



Barometric pressure compensation	Automatic		
Zero Calibration	Support		
Calibration	Support		
Apnea alarm delay	10 s, 15 s, 20 s, 25 s, 30	s, 35 s, 40 s	
Safety Specifications			
	IEC 60601-1: 2005+A1 :2012; IEC 60601-1-2: 2014;		
Compliant with Standards	EN 60601-1: 2006+A1 :2013; EN 60601-1-2: 2015;		
	IEC 60601-2-49: 2011		
Anti-electroshock Type	Class I equipment and internal powered equipment		
Anti-electroshock Degree	CF		
Ingress Protection	IPX1		
Environmental Specifications			
Tomporoturo	Working	+0°C to +40°C (32 °F ~104 °F)	
Temperature	Transport and Storage	-20°C to +55°C (-4 °F ~131 °F)	
الليسم مانان	Working	15%RH to 95%RH (non-condensing)	
Humidity	Transport and Storage	15%RH to 95%RH (non-condensing)	
Altitudo	Working	86 kPa to 106 kPa	
Altitude	Transport and Storage	70 kPa to 106 kPa	

<sup>\*</sup> Specifications are subject to change without prior notice

## ONTROLab.

